



HIC products are proudly distributed by

NUCMEDCOR

P.O. Box 29475
San Francisco, CA 94129

Huayi Isotopes Co.

The leading manufacturer of Oxygen 18 Water & PET Precursors

FILL-EASE™

Sterile Vial


According to cGMP Requirement

Catalog No.: SV-100B
Lot No.: SV-100B-200501

Date of Manufacturing: May. 2020
Expiry Date: May. 2022

Product Description

Unit: mm



Filling Volume: 107.0ml ± 2.0ml
 Height: 95.0mm ± 1.0mm
 Thickness of bottle-lip: 4.0mm ± 0.3mm
 Height of bottle-neck: 3.5mm ± 0.3mm
 Effective height of bottle-body: 62.0mm ± 2.0mm
 Outer diameter of bottle-mouth: 19.8mm ± 0.2mm
 Inner diameter of bottle-mouth: 12.6mm ± 0.2mm
 Outer diameter of bottle-neck: 16.7mm ± 0.5mm
 Body diameter: 51.5mm ± 0.6mm
 Material: Neutral borosilicate glass
 (PB-053)
 Sealing type: Chlorobutyl rubber stopper
 (PC-003)
 Cap: 20mm aluminum cap
 (PA-013)

Note

- The product should be stored at room temperature.
- Manufactured according to requirements of cGMP, ISO 9001 and ISO 13485.
- We certify that all test results meets the acceptable criteria of Huayi specification current QS-SV.
- Tested for sterility and Bacterial endotoxins.

Drafted

Xingqi Huang

Date: 2020.06.13

Reviewed

Siqing Lv

Date: 2020.06.13

Approved

Jie Shen

Date: 2020.06.13



Huayi Isotopes Co.

The leading manufacturer of Oxygen 18 Water & PET Precursors

FILL-EASE™

Sterile Vial

According to cGMP Requirement

Catalog No.: SV-100B

Lot No.: SV-100B-200501

Date of Manufacturing: May. 2020

Expiry Date: May. 2022

Certificate of Analysis

Step/Procedure	Method	Specification	Result
Materials			
Vial	Current USP<660> Current EP<3.2.1>	Neutral borosilicate glass; PB-053	Lot No.: 2003102
Stopper	Current USP<381> Current EP<3.2.9>	Chlorobutyl rubber stopper, no latex; PC-003	Lot No.: 2003013
Cap	In house	20mm aluminum cap; PA-013	Lot No.: 1910037
Appearance	Visual inspection	Integrity ,no obvious spots and scratches	Integrity ,no obvious spots and scratches
Test for sterility	Current USP<71> Current EP<2.6.1>	Passed test	Passed test
Bacterial Endotoxin	Current USP<85> Current EP<2.6.14>	≤ 0.25 EU/mL	Conforms
Particles, ≥25 μm	Current USP<788> Current EP<2.9.19>	< 150	41
Particles, ≥10 μm	Current USP<788> Current EP<2.9.19>	< 1500	65
Sealing Capacity	In house	No leak	No leak